

# Return & Repair Form:

## Customer Information:

Company Name: Hyndhurst Police B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 24 May 11 Date Given to Service: 24 May 11  
 Carrier: PedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

## Product Information:

Product: 7410-~~7110~~-6510-6810-7510  
 9510- Drug Tester 5000  
 Description: A - B - Plus - Screener - Demo  
 Serial #: ARXD - 0019  
 Printer Ser #: AR  
 Sim Ser #: \_\_\_\_\_  
 Probe Ser # DD P  
 Whole Instrument  Top 1/2  
 Other \_\_\_\_\_  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test #

Reason for Return:

Part Number	Description	Qty	Total Cost
<u>0808455</u>	<u>FC</u>	<u>1</u>	<u>n/cw</u>
<u>mp cal 71</u>	<u>cal</u>	<u>1</u>	<u>n/cw</u>
<u>mp laba</u>	<u>labor</u>	<u>.5</u>	<u>n/cw</u>

Repair Notes:

ECT# ARCA-0796

Install FC

Cal w/DC; eps check

Service Technician: [Signature]

Date: 5/27/2011