

# RETURN AND REPAIR FORM



## Customer Information

B:

S:

Company Name: Essex County Sheriff's Office, Newark, NJ

Date Received: 02/25/2016 Date given to service: 02/25/2016

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 Whole

Serial Number: AR XE-0069  
 Printer Serial#: AR  
 Sim Serial#: DD  
 Probe Serial#: DD

Warranty Expires: December 2013 (47032)

Description:  A  B  Plus  Demo  Screener  Trade In

## Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

## Repair Information:

## Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	
4414166	Labor	.5	
6808455	Fuel Cell	1	
6808486	Spacer Plates	2	

Repair Notes: Replaced defective fuel cell and spacer plates.

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

Fuel Cell = ARJB-1468  
 CAL W/QC AND OPS CHECK

Service Technician BC Date: 04/14/2016