

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Moonachie P.D., Moonachie, NJ

Date Received: 11/14/2016

Date given to service: 11/14/2016

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR XE-0072

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: December 2016

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

| Part Number | Description | Qty | Total Cost |
|-------------|---------------|-----|------------|
| 4414161 | Calibration | 1 | NC/W |
| 4414166 | Labor | 1.5 | NC/W |
| 6808455 | Fuel Cell | 1 | NC/W |
| 6808486 | Spacer Plates | 2 | NC/W |
| 8312820 | Pump | 1 | NC/W |
| | | | |
| | | | |
| | | | |

Repair Notes: Replaced defective fuel cell, pump, and spacer plates.

Fuel Cell = ARJE-1951

CAL W/QC AND OPS CHECK

Service Technician: BC

Date: 11/16/2016