

# Return & Repair Form:

## Customer Information:

Company Name: Saddle River P.D. B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 7-1-09 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

## Product Information:

Product: 7410 - 7110 - 6510 - 6810 Serial #: AR 22-0134  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -  
 Whole Instrument Top 1/2 Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P -  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test #

Reason for Return: NEW Cuvette

| Part Number      | Description             | Qty       | Total Cost      |
|------------------|-------------------------|-----------|-----------------|
| <u>MP Cal 71</u> | <u>Cal</u>              | <u>1</u>  | <u>N/C WARR</u> |
| <u>6809512</u>   | <u>Grammar / Vent-1</u> | <u>1</u>  | <u>N/C WARR</u> |
| <u>6809500</u>   | <u>Cuvette</u>          | <u>1</u>  | <u>N/C WARR</u> |
|                  |                         |           |                 |
|                  |                         |           |                 |
|                  |                         |           |                 |
| <u>MP Labor</u>  | <u>Labor</u>            | <u>.5</u> | <u>N/C WARR</u> |

Repair Notes:  
Replace Cuvette  
Cal w/Qc; Ops Check  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

Service Technician: AS Date: 7-8-09