

Return & Repair Form:

Customer Information:

Company Name: Edgewater B: _____ S: _____
 Date Received: 12-11-09 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410-7110-6510-6810-7510 Serial #: AR 26-0132
 9510- Drug Tester 5000
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top ½ Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: FL / IR Detektor

| Part Number | Description | Qty | Total Cost |
|------------------|--------------------|-----------|----------------|
| <u>0.809458</u> | <u>IR Detektor</u> | <u>1</u> | <u>N/C WAW</u> |
| <u>MP Cal 71</u> | <u>Cal</u> | <u>1</u> | <u>N/C WAW</u> |
| <u>MP Labor</u> | <u>Labor</u> | <u>.5</u> | <u>N/C WAW</u> |

Repair Notes:

NEW IR detektor
Clean Cuvette
Cal w/QC ; OPS check

Service Technician: JS

Date: 12/16/09