

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Edgewater P.D. Edgewater, NJ

Date Received: 08/18/11 Date given to service: _____

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR ZL-0137

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
<u>MP Cal 71</u>	<u>Calibrate</u>	<u>1</u>	<u>N/CW</u>
<u>6808455</u>	<u>Fuel Cell</u>	<u>1</u>	<u>N/CW</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/CW</u>

Repair Notes: Clean spots off cuvette, installed new fuel cell # ARCH-1155.

Cal w/DC & ops check

Service Technician: [Signature] Date: 8-23-11