

# RETURN AND REPAIR FORM



## Customer Information

B:

S:

Company Name: Montvale P.D., Montvale, NJ

Date Received: 09/30/13

Date given to service: 09/30/13

Carrier:  FedEx  UPS  USPS

Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 Whole

Serial Number: AR ZL-0142

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: January 2014

Description:  A  B  Plus  Demo  Screener  Trade In

## Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

## Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	NC/W
MP Labor	Labor	9	NC/W
8315075	Alcotest 7110 Motherboard	1	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plates	2	NC/W

Repair Notes: Replaced defective motherboard, fuel cell, and spacer plates.

Motherboard was changed due to hardware issue, was not letting the 0.400 to be ran would give BLOWING NOT ALLOWED error.

Fuel Cell = AREF-1808

CAL W/QC AND OPS CHECK

Service Technician: BC

Date: 10/18/13