RETURN AND REPAIR FORM

Dräger

Customer Information

B: S:

Company Name: Wall Twp. P.D., Wall, NJ

Date Received: 03/18/14 **Date given to service:** 03/18/14

Carrier: X FedEx UPS USPS **Shipping Method:** GRD X 3DAY 2DAY NDA-PRI NDA-STD

Product:

	6510	6810
X	7110	8610
	7510	DT5000
	7410 Upper-half	9510
	7410 Whole	

Serial Number: AR ZL-0146

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires:

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter	Regulator	Printer Ribbon	Printer Paper
Mouthpieces	9510 Stylus	9510 Top Cover	Carrying Case
Dry Gas	Other (please specify)		

Repair Information

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC/W
MP Labor	Labor	10	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plates	2	NC/W
8315075	Alcotest 7110 Motherboard	1	NC/W
6809458	IR Detector	1	NC/W

Repair Notes: Replaced defective Alcotest 7110 Motherboard, IR Detector, fuel cell and spacer plates. Alcotest 7110 Motherboard was replaced due to the IRV not stablizing and the IR Detector was replaced due to unit giving "BLOWING NOT ALLOWED" error.

Fuel Cell = ARFB-0252

CAL W/QC AND OPS CHECK

Service Technician BC **Date:** 04/08/14

sns\xfs010\Public\Forms\return and repair form 07_21_2011.pdf